

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021396.D
 Acq On : 23 Mar 2021 19:17
 Operator : JC/MD
 Sample : M1724-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 26207

Quant Time: Mar 24 01:59:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77178	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	126721	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	114850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	61714	48.75	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 97.50%			
35) Dibromofluoromethane	5.464	113	43723	50.71	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.42%			
50) Toluene-d8	8.694	98	158164	50.55	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 101.10%			
62) 4-Bromofluorobenzene	11.118	95	54971	45.24	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 90.48%			
Target Compounds						
11) Tert butyl alcohol	3.005	59	1325	6.23	ug/l #	88
16) Acetone	2.417	43	10503	22.55	ug/l #	82
20) Methylene Chloride	2.834	84	6402	6.68	ug/l #	74
25) 2-Butanone	4.635	43	4024	5.62	ug/l #	85
43) Isopropyl Acetate	6.405	43	9192	3.79	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021396.D
Acq On : 23 Mar 2021 19:17
Operator : JC/MD
Sample : M1724-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
26207

Quant Time: Mar 24 01:59:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

