

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021401.D
 Acq On : 23 Mar 2021 22:24
 Operator : JC/MD
 Sample : VX0323WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0323WBL02

Quant Time: Mar 24 02:02:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	76437	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	127295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	112714	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	48214	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	61407	48.98	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.96%
35) Dibromofluoromethane	5.464	113	43686	50.43	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.86%
50) Toluene-d8	8.694	98	155732	49.55	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.10%
62) 4-Bromofluorobenzene	11.118	95	55441	45.42	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.84%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021401.D
Acq On : 23 Mar 2021 22:24
Operator : JC/MD
Sample : VX0323WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0323WBL02

Quant Time: Mar 24 02:02:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

