

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032321\  
 Data File : VX021405.D  
 Acq On : 23 Mar 2021 23:57  
 Operator : JC/MD  
 Sample : M1735-09 20X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 36 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D7-20210315

Quant Time: Mar 24 03:33:28 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031621W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Mar 16 14:00:21 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min)  |
|------------------------------|--------|-------|----------|----------|--------|-----------|
| -----                        |        |       |          |          |        |           |
| Internal Standards           |        |       |          |          |        |           |
| 1) Pentafluorobenzene        | 5.629  | 168   | 77082    | 50.00    | ug/l   | 0.00      |
| 34) 1,4-Difluorobenzene      | 6.829  | 114   | 125047   | 50.00    | ug/l   | 0.00      |
| 63) Chlorobenzene-d5         | 10.094 | 117   | 109066   | 50.00    | ug/l   | 0.00      |
| 72) 1,4-Dichlorobenzene-d4   | 12.059 | 152   | 44016    | 50.00    | ug/l   | 0.00      |
| System Monitoring Compounds  |        |       |          |          |        |           |
| 33) 1,2-Dichloroethane-d4    | 6.029  | 65    | 60564    | 47.90    | ug/l   | 0.00      |
| Spiked Amount                | 50.000 | Range | 78 - 117 | Recovery | =      | 95.80%    |
| 35) Dibromofluoromethane     | 5.464  | 113   | 43463    | 51.08    | ug/l   | 0.00      |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =      | 102.16%   |
| 50) Toluene-d8               | 8.694  | 98    | 152938   | 49.54    | ug/l   | 0.00      |
| Spiked Amount                | 50.000 | Range | 92 - 112 | Recovery | =      | 99.08%    |
| 62) 4-Bromofluorobenzene     | 11.118 | 95    | 53807    | 44.88    | ug/l   | 0.00      |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =      | 89.76%    |
| Target Compounds             |        |       |          |          |        |           |
| 9) 1,1,2-Trichlorotrifluo... | 2.364  | 101   | 1081     | 1.34     | ug/l   | Qvalue 94 |
| 12) 1,1-Dichloroethene       | 2.352  | 96    | 312      | 0.40     | ug/l # | 80        |
| 44) Trichloroethene          | 7.188  | 130   | 68760    | 74.55    | ug/l   | 96        |
| -----                        |        |       |          |          |        |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032321\  
Data File : VX021405.D  
Acq On : 23 Mar 2021 23:57  
Operator : JC/MD  
Sample : M1735-09 20X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 36 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D7-20210315

Quant Time: Mar 24 03:33:28 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031621W.M  
Quant Title : SW846 8260  
QLast Update : Tue Mar 16 14:00:21 2021  
Response via : Initial Calibration

