

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021411.D
 Acq On : 24 Mar 2021 02:18
 Operator : JC/MD
 Sample : M1770-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105D

Quant Time: Mar 24 03:35:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	72987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	122513	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	109731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	46721	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	59733	49.90	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.80%
35) Dibromofluoromethane	5.458	113	42159	50.57	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.14%
50) Toluene-d8	8.694	98	148685	49.16	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.32%
62) 4-Bromofluorobenzene	11.118	95	55027	46.84	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.68%
Target Compounds						
16) Acetone	2.417	43	2187	4.97	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021411.D
Acq On : 24 Mar 2021 02:18
Operator : JC/MD
Sample : M1770-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105D

Quant Time: Mar 24 03:35:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

