

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021412.D
 Acq On : 24 Mar 2021 02:41
 Operator : JC/MD
 Sample : M1770-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-105I

Quant Time: Mar 24 03:35:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.634	168	72346	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	122475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	101976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	40701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	60650	51.11	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.22%		
35) Dibromofluoromethane	5.470	113	43403	52.08	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.16%		
50) Toluene-d8	8.694	98	148132	48.99	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 97.98%		
62) 4-Bromofluorobenzene	11.118	95	49126	41.83	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 83.66%		
Target Compounds						
16) Acetone	2.422	43	2000	4.58	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021412.D
Acq On : 24 Mar 2021 02:41
Operator : JC/MD
Sample : M1770-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-105I

Quant Time: Mar 24 03:35:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

