

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
 Data File : VX021420.D
 Acq On : 24 Mar 2021 05:49
 Operator : JC/MD
 Sample : M1750-02 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1015

Quant Time: Mar 24 06:10:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 16 14:00:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	77925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	129814	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	111862	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	45208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	62429	48.84	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.68%
35) Dibromofluoromethane	5.464	113	44634	50.53	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.06%
50) Toluene-d8	8.694	98	158352	49.41	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.82%
62) 4-Bromofluorobenzene	11.118	95	53462	42.95	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.90%
Target Compounds						
16) Acetone	2.417	43	82776	176.02	ug/l #	87
52) Toluene	8.765	92	6096	2.73	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032321\
Data File : VX021420.D
Acq On : 24 Mar 2021 05:49
Operator : JC/MD
Sample : M1750-02 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1015

Quant Time: Mar 24 06:10:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031621W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 16 14:00:21 2021
Response via : Initial Calibration

