

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027663.D
 Acq On : 23 Mar 2022 13:29
 Operator : JC/MD
 Sample : N1998-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB031622

Quant Time: Mar 24 04:35:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	351640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	563190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	526801	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	241448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206501	47.670	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.340%
35) Dibromofluoromethane	5.391	113	183119	50.537	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.080%
50) Toluene-d8	8.653	98	650920	47.472	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	11.079	95	246502	48.911	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.820%
Target Compounds						
16) Acetone	2.380	43	4405	2.495	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027663.D
Acq On : 23 Mar 2022 13:29
Operator : JC/MD
Sample : N1998-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB031622

Quant Time: Mar 24 04:35:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

