

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032322\  
 Data File : VX027664.D  
 Acq On : 23 Mar 2022 13:53  
 Operator : JC/MD  
 Sample : N2028-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-114D

Quant Time: Mar 24 04:35:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031722W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Mar 17 15:02:59 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.556	168	381412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	608505	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	538578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	244739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	203299	43.268	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.540%
35) Dibromofluoromethane	5.391	113	196021	50.069	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.140%
50) Toluene-d8	8.653	98	687798	46.426	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.860%
62) 4-Bromofluorobenzene	11.079	95	249627	45.843	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.680%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5507	2.875	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	5038	0.422	ug/l	98
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX032322\  
 Data File : VX027664.D  
 Acq On : 23 Mar 2022 13:53  
 Operator : JC/MD  
 Sample : N2028-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MW-114D

Quant Time: Mar 24 04:35:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031722W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Mar 17 15:02:59 2022  
 Response via : Initial Calibration

