

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027665.D
 Acq On : 23 Mar 2022 14:16
 Operator : JC/MD
 Sample : N2028-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-116D

Quant Time: Mar 24 04:35:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	333939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	536773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	506430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	220329	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	196798	47.839	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.680%
35) Dibromofluoromethane	5.391	113	175325	50.767	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.540%
50) Toluene-d8	8.653	98	629869	48.198	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.400%
62) 4-Bromofluorobenzene	11.079	95	230266	47.938	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3580	2.135	ug/l	95
19) Methyl tert-butyl Ether	3.117	73	7162	0.685	ug/l	94
67) Ethyl Benzene	10.195	91	6277	0.352	ug/l #	87
68) m/p-Xylenes	10.305	106	2520	0.359	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	3393	0.252	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027665.D
Acq On : 23 Mar 2022 14:16
Operator : JC/MD
Sample : N2028-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-116D

Quant Time: Mar 24 04:35:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

