

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027668.D
 Acq On : 23 Mar 2022 15:27
 Operator : JC/MD
 Sample : N1908-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-8-9-3

Quant Time: Mar 23 17:27:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	350953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	570213	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	537890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	240054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	208600	48.249	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.500%
35) Dibromofluoromethane	5.391	113	187618	51.141	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.280%
50) Toluene-d8	8.653	98	696802	50.193	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.380%
62) 4-Bromofluorobenzene	11.079	95	275483	53.989	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30564	17.344	ug/l	94
18) Methyl Acetate	2.709	43	21789	5.159	ug/l	94
20) Methylene Chloride	2.788	84	11111	2.933	ug/l	98
25) 2-Butanone	4.568	43	19444	7.334	ug/l	95
43) Isopropyl Acetate	6.349	43	27046	3.364	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027668.D
Acq On : 23 Mar 2022 15:27
Operator : JC/MD
Sample : N1908-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-8-9-3

Quant Time: Mar 23 17:27:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

