

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027674.D
 Acq On : 23 Mar 2022 17:48
 Operator : JC/MD
 Sample : N1905-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14A-1A-4

Quant Time: Mar 23 18:38:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	354675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	587729	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	532178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	249503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	208671	47.759	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.520%
35) Dibromofluoromethane	5.391	113	188220	49.776	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.560%
50) Toluene-d8	8.653	98	683533	47.769	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.540%
62) 4-Bromofluorobenzene	11.079	95	248813	47.309	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.620%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	47873	26.881	ug/l	94
20) Methylene Chloride	2.788	84	10526	2.749	ug/l	88
25) 2-Butanone	4.562	43	26964	10.063	ug/l	98
39) Methylcyclohexane	7.379	83	35148	5.409	ug/l	92
40) Benzene	6.044	78	22726	1.533	ug/l	96
43) Isopropyl Acetate	6.348	43	28663	3.459	ug/l	96
68) m/p-Xylenes	10.305	106	10984	1.489	ug/l	98
73) Isopropylbenzene	10.964	105	79529	4.376	ug/l	99
78) n-propylbenzene	11.305	91	63921	3.204	ug/l	100
85) sec-Butylbenzene	11.890	105	25916	1.415	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027674.D
Acq On : 23 Mar 2022 17:48
Operator : JC/MD
Sample : N1905-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14A-1A-4

Quant Time: Mar 23 18:38:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

