

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027677.D
 Acq On : 23 Mar 2022 19:22
 Operator : JC/MD
 Sample : N2028-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB031822

Quant Time: Mar 24 04:36:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	348664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	564262	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	533344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236431	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	203315	47.335	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.680%
35) Dibromofluoromethane	5.391	113	183961	50.673	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.340%
50) Toluene-d8	8.652	98	669045	48.701	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.085	95	239142	47.361	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027677.D
Acq On : 23 Mar 2022 19:22
Operator : JC/MD
Sample : N2028-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB031822

Quant Time: Mar 24 04:36:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

