

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027678.D
 Acq On : 23 Mar 2022 19:45
 Operator : JC/MD
 Sample : N2028-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Mar 24 04:37:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	341830	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	557615	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	493140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	224617	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	201227	47.79	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.58%
35) Dibromofluoromethane	5.391	113	181500	50.59	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.18%
50) Toluene-d8	8.653	98	645404	47.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.08%
62) 4-Bromofluorobenzene	11.085	95	227313	45.55	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.10%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027678.D
Acq On : 23 Mar 2022 19:45
Operator : JC/MD
Sample : N2028-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Mar 24 04:37:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

