

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
 Data File : VX027680.D
 Acq On : 23 Mar 2022 20:32
 Operator : JC/MD
 Sample : N2028-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-116

Quant Time: Mar 24 04:37:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 17 15:02:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	551235	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	522719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	200476	47.842	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.680%
35) Dibromofluoromethane	5.391	113	179602	50.641	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.280%
50) Toluene-d8	8.653	98	643866	47.976	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.079	95	230700	46.769	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	5879	3.442	ug/l #	87
19) Methyl tert-butyl Ether	3.123	73	4083	0.383	ug/l #	82
30) Chloroform	5.099	83	9439	1.458	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032322\
Data File : VX027680.D
Acq On : 23 Mar 2022 20:32
Operator : JC/MD
Sample : N2028-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-116

Quant Time: Mar 24 04:37:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031722W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 17 15:02:59 2022
Response via : Initial Calibration

