

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032419\
 Data File : VX008381.D
 Acq On : 24 Mar 2019 11:02
 Operator : JC/SP
 Sample : K1969-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB06-20190315

Quant Time: Mar 25 06:37:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	382377	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	328231	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	164636	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	5.49	113	120013	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
50) Toluene-d8	8.71	98	481884	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	159710	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	
Target Compounds						
17) Carbon Disulfide	2.57	76	1969	0.242	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032419\
Data File : VX008381.D
Acq On : 24 Mar 2019 11:02
Operator : JC/SP
Sample : K1969-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB06-20190315

Quant Time: Mar 25 06:37:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

