

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032419\
 Data File : VX008385.D
 Acq On : 24 Mar 2019 12:36
 Operator : JC/SP
 Sample : K2059-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB032019

Quant Time: Mar 25 06:44:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	224569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	392289	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138633	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168463	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
35) Dibromofluoromethane	5.50	113	124338	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	493962	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.13	95	164687	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
Target Compounds						
16) Acetone	2.44	43	4213	2.461	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032419\
Data File : VX008385.D
Acq On : 24 Mar 2019 12:36
Operator : JC/SP
Sample : K2059-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB032019

Quant Time: Mar 25 06:44:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

