

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032419\
 Data File : VX008401.D
 Acq On : 24 Mar 2019 18:48
 Operator : JC/SP
 Sample : K1945-12 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M4-T4-UV(POST-HYPO)

Quant Time: Mar 25 07:58:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	238277	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	402061	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	361935	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152372	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	167841	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
35) Dibromofluoromethane	5.49	113	125928	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
50) Toluene-d8	8.71	98	520953	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	180264	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032419\
Data File : VX008401.D
Acq On : 24 Mar 2019 18:48
Operator : JC/SP
Sample : K1945-12 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M4-T4-UV(POST-HYPO)

Quant Time: Mar 25 07:58:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration

