

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032419\
 Data File : VX008400.D
 Acq On : 24 Mar 2019 18:24
 Operator : JC/SP
 Sample : K1945-06 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M4-T4-OZ(POST-HYPO)

Quant Time: Mar 25 07:57:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 15:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	234383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	399905	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	360181	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149407	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	167049	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
35) Dibromofluoromethane	5.49	113	124662	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
50) Toluene-d8	8.71	98	515732	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
62) 4-Bromofluorobenzene	11.13	95	178981	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032419\
Data File : VX008400.D
Acq On : 24 Mar 2019 18:24
Operator : JC/SP
Sample : K1945-06 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M4-T4-OZ(POST-HYPO)

Quant Time: Mar 25 07:57:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 15:27:34 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX008400.D

Acq: 24 Mar 2019 18:24

Instrument :

MSVOA_X

ClientSampleId :


M4-T4-OZ(POST-HYPO)

Tot Ion:168 Resp: 234383

Ion Ratio Lower Upper




168 100

99 61.6 37.6 56.4#



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 46.729 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX008400.D


Acq: 24 Mar 2019 18:24

Tgt Ion: 65 Resp: 167049

Ion Ratio Lower Upper



65 100


67 53.8 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 935
 Delta R.T. -0.01 min
 Lab File: VX008400.D
 Acq: 24 Mar 2019 18:24




Instrument :
 MSVOA_X
 ClientSampleId :
 M4-T4-OZ(POST-HYPO)


Tot Ion	Ratio	Lower	Upper
114	100		
63	24.3	0.0	39.8
88	17.2	0.0	31.4



#35
 Dibromofluoromethane
 Concen: 48.114 ug/l
 RT: 5.49 min Scan# 712
 Delta R.T. 0.00 min
 Lab File: VX008400.D
 Acq: 24 Mar 2019 18:24

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.4	81.4	122.0
192	19.4	19.4	29.0









#50
Toluene-d8
Concen: 50.325 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX008400.D
Acq: 24 Mar 2019 18:24




Instrument :
MSVOA_X
ClientSampleId :
M4-T4-OZ(POST-HYPO)


Tot Ion: 98 Resp: 515732
Ion Ratio Lower Upper
98 100
100 63.8 51.7 77.5



#62
4-Bromofluorobenzene
Concen: 47.568 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX008400.D
Acq: 24 Mar 2019 18:24

Tgt Ion: 95 Resp: 178981
Ion Ratio Lower Upper
95 100
174 73.7 0.0 182.8
176 71.2 0.0 178.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. 0.00 min

Lab File: VX008400.D

Acq: 24 Mar 2019 18:24

Instrument :

MSVOA_X

ClientSampleId :

M4-T4-OZ(POST-HYPO)


Tot Ion:117 Resp: 360181

Ion Ratio Lower Upper

117 100

82 62.6 44.3 66.5




119 32.1 25.3 37.9



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX008400.D

Acq: 24 Mar 2019 18:24


Tgt Ion:152 Resp: 149407

Ion Ratio Lower Upper

152 100

115 61.2 38.6 115.7

150 156.4 0.0 348.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

