


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX032419\
Data File : VX008377.D
Acq On : 24 Mar 2019 09:14
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032319W.M
Title : SW846 8260
Last Update : Sat Mar 23 15:27:34 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	10501	PASS
75	95	30	60	52.8	25114	PASS
95	95	100	100	100.0	47589	PASS
96	95	5	9	6.7	3200	PASS
173	174	0.00	2	1.3	445	PASS
174	95	50	100	72.1	34304	PASS
175	174	5	9	7.4	2555	PASS
176	174	95	101	97.8	33560	PASS
177	176	5	9	7.0	2336	PASS