

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021432.D
 Acq On : 24 Mar 2021 13:51
 Operator : JC/MD
 Sample : VX0324WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0324WBL01

Quant Time: Mar 24 17:17:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	62183	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	112068	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	117615	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	57466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.023	65	52424	50.67	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.34%
35) Dibromofluoromethane	5.464	113	39657	49.82	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.64%
50) Toluene-d8	8.694	98	142116	50.10	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.20%
62) 4-Bromofluorobenzene	11.118	95	62402	56.58	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.16%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021432.D
Acq On : 24 Mar 2021 13:51
Operator : JC/MD
Sample : VX0324WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0324WBL01

Quant Time: Mar 24 17:17:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

