

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021438.D
 Acq On : 24 Mar 2021 16:10
 Operator : JC/MD
 Sample : M1735-10 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D6-20210315

Quant Time: Mar 25 02:54:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	67494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	122950	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	125079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	56188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	56789	50.57	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.14%
35) Dibromofluoromethane	5.470	113	42707	48.90	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.80%
50) Toluene-d8	8.694	98	155296	49.90	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.118	95	65470	54.11	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.22%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.364	101	1649	2.37	ug/l	97
44) Trichloroethene	7.188	130	111768	123.81	ug/l	91
64) Tetrachloroethene	9.318	164	232	0.28	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021438.D
Acq On : 24 Mar 2021 16:10
Operator : JC/MD
Sample : M1735-10 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D6-20210315

Quant Time: Mar 25 02:54:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

