

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021445.D
 Acq On : 24 Mar 2021 18:46
 Operator : JC/MD
 Sample : M1774-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30808

Quant Time: Mar 25 02:54:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	66854	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.829	114	123665	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	123124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	55861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	56963	51.21	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.42%
35) Dibromofluoromethane	5.470	113	43253	49.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.48%
50) Toluene-d8	8.694	98	155048	49.53	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.06%
62) 4-Bromofluorobenzene	11.118	95	62184	51.10	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.20%
Target Compounds						
11) Tert butyl alcohol	3.017	59	1682	9.15	ug/l	Qvalue 99
16) Acetone	2.423	43	25510	60.65	ug/l	# 88
25) 2-Butanone	4.641	43	4177	6.55	ug/l	# 88
52) Toluene	8.765	92	2094	0.99	ug/l	99
95) Naphthalene	13.818	128	4168	1.14	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021445.D
Acq On : 24 Mar 2021 18:46
Operator : JC/MD
Sample : M1774-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30808

Quant Time: Mar 25 02:54:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

