

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021454.D
 Acq On : 24 Mar 2021 23:27
 Operator : JC/MD
 Sample : M1787-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02-20210318

Quant Time: Mar 25 06:52:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	64665	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	115783	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115474	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	54065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53309	49.55	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.10%
35) Dibromofluoromethane	5.458	113	40564	49.32	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.64%
50) Toluene-d8	8.694	98	144058	49.15	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.30%
62) 4-Bromofluorobenzene	11.118	95	58782	51.59	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.18%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021454.D
Acq On : 24 Mar 2021 23:27
Operator : JC/MD
Sample : M1787-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02-20210318

Quant Time: Mar 25 06:52:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

