

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021456.D
 Acq On : 25 Mar 2021 00:14
 Operator : JC/MD
 Sample : M1770-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-106D

Quant Time: Mar 25 06:53:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	63699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	113743	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	116029	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	50608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	53742	50.71	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.42%	
35) Dibromofluoromethane	5.464	113	40763	50.45	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.90%	
50) Toluene-d8	8.694	98	145296	50.47	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.94%	
62) 4-Bromofluorobenzene	11.118	95	58296	52.08	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.16%	
Target Compounds						
						Qvalue
16) Acetone	2.422	43	3129	7.81	ug/l	92
51) 4-Methyl-2-Pentanone	8.617	43	3931	3.05	ug/l	# 84
68) m/p-Xylenes	10.335	106	1067	0.70	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021456.D
Acq On : 25 Mar 2021 00:14
Operator : JC/MD
Sample : M1770-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-106D

Quant Time: Mar 25 06:53:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

