

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021457.D
 Acq On : 25 Mar 2021 00:38
 Operator : JC/MD
 Sample : M1770-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-104D

Quant Time: Mar 25 06:53:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	62991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	115679	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	115508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	54755	52.24	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.48%
35) Dibromofluoromethane	5.470	113	40518	49.31	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.62%
50) Toluene-d8	8.694	98	144579	49.38	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.76%
62) 4-Bromofluorobenzene	11.118	95	58472	51.36	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.72%
Target Compounds						
16) Acetone	2.423	43	2088	5.27	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021457.D
Acq On : 25 Mar 2021 00:38
Operator : JC/MD
Sample : M1770-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-104D

Quant Time: Mar 25 06:53:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

