

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021458.D
 Acq On : 25 Mar 2021 01:01
 Operator : JC/MD
 Sample : M1770-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-104I

Quant Time: Mar 25 06:53:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	64531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	116603	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	119137	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52387	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	55167	51.38	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.76%		
35) Dibromofluoromethane	5.470	113	41373	49.95	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.90%		
50) Toluene-d8	8.694	98	150725	51.07	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.14%		
62) 4-Bromofluorobenzene	11.118	95	60036	52.32	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.64%		
Target Compounds						Qvalue
16) Acetone	2.423	43	3254	8.02	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021458.D
Acq On : 25 Mar 2021 01:01
Operator : JC/MD
Sample : M1770-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-104I

Quant Time: Mar 25 06:53:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

