

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021461.D
 Acq On : 25 Mar 2021 02:12
 Operator : JC/MD
 Sample : M1770-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-101S

Quant Time: Mar 25 06:54:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	65946	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	121750	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	123866	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	54424	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	56315	51.32	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.64%	
35) Dibromofluoromethane	5.464	113	42746	49.43	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.86%	
50) Toluene-d8	8.694	98	153340	49.76	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.52%	
62) 4-Bromofluorobenzene	11.118	95	63085	52.65	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.30%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.352	96	4533	6.82	ug/l #	81
16) Acetone	2.422	43	1351	3.26	ug/l #	83
24) 1,1-Dichloroethane	3.670	63	8306	5.88	ug/l	97
27) cis-1,2-Dichloroethene	4.570	96	5300	6.25	ug/l	96
30) Chloroform	5.170	83	2573	1.73	ug/l	89
32) 1,1,1-Trichloroethane	5.464	97	1432	1.08	ug/l #	45
44) Trichloroethene	7.188	130	19521	21.84	ug/l	90
64) Tetrachloroethene	9.318	164	8875	10.70	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021461.D
Acq On : 25 Mar 2021 02:12
Operator : JC/MD
Sample : M1770-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-101S

Quant Time: Mar 25 06:54:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

