

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021464.D
 Acq On : 25 Mar 2021 03:22
 Operator : JC/MD
 Sample : M1787-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D2-20210318

Quant Time: Mar 25 06:55:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	63366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	118607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	119401	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	52576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	54554	51.74	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.48%		
35) Dibromofluoromethane	5.464	113	42877	50.89	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.78%		
50) Toluene-d8	8.694	98	149347	49.75	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.50%		
62) 4-Bromofluorobenzene	11.118	95	61008	52.27	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.54%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.364	101	23461	35.97	ug/l	96
12) 1,1-Dichloroethene	2.352	96	13884	21.74	ug/l	89
16) Acetone	2.417	43	1432	3.59	ug/l #	79
24) 1,1-Dichloroethane	3.670	63	3283	2.42	ug/l #	91
27) cis-1,2-Dichloroethene	4.564	96	2561	3.14	ug/l	98
30) Chloroform	5.170	83	1464	1.02	ug/l #	81
32) 1,1,1-Trichloroethane	5.464	97	3041	2.39	ug/l #	93
38) Carbon Tetrachloride	5.758	117	2460	1.97	ug/l #	94
44) Trichloroethene	7.188	130	263727	302.83	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021464.D
Acq On : 25 Mar 2021 03:22
Operator : JC/MD
Sample : M1787-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D2-20210318

Quant Time: Mar 25 06:55:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

