

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021468.D
 Acq On : 25 Mar 2021 04:56
 Operator : JC/MD
 Sample : M1787-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-125D2-20210319

Quant Time: Mar 25 06:56:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	61545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	111500	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110073	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	47722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	52902	51.66	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.32%	
35) Dibromofluoromethane	5.470	113	40153	50.70	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.40%	
50) Toluene-d8	8.694	98	141437	50.11	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.22%	
62) 4-Bromofluorobenzene	11.118	95	55287	50.39	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.78%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.370	101	14985	23.65	ug/l	95
12) 1,1-Dichloroethene	2.358	96	3614	5.83	ug/l #	75
27) cis-1,2-Dichloroethene	4.564	96	4561	5.77	ug/l	98
44) Trichloroethene	7.188	130	300231	366.72	ug/l	95
64) Tetrachloroethene	9.318	164	4861	6.59	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021468.D
Acq On : 25 Mar 2021 04:56
Operator : JC/MD
Sample : M1787-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-125D2-20210319

Quant Time: Mar 25 06:56:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

