

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021469.D
 Acq On : 25 Mar 2021 05:19
 Operator : JC/MD
 Sample : M1787-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-125D3-20210319

Quant Time: Mar 25 06:57:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	59117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	107326	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	110176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	47992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	51368	52.22	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.44%
35) Dibromofluoromethane	5.464	113	39414	51.70	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.40%
50) Toluene-d8	8.694	98	138002	50.80	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.60%
62) 4-Bromofluorobenzene	11.118	95	54203	51.32	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.64%
Target Compounds						
					Qvalue	
3) Chloromethane	1.316	50	1753	2.39	ug/l	# 87
9) 1,1,2-Trichlorotrifluo...	2.364	101	59834	98.33	ug/l	97
27) cis-1,2-Dichloroethene	4.564	96	2126	2.80	ug/l	96
44) Trichloroethene	7.188	130	160774	204.02	ug/l	92
64) Tetrachloroethene	9.317	164	5132	6.95	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021469.D
Acq On : 25 Mar 2021 05:19
Operator : JC/MD
Sample : M1787-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-125D3-20210319

Quant Time: Mar 25 06:57:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

