

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
 Data File : VX021471.D
 Acq On : 25 Mar 2021 06:06
 Operator : JC/MD
 Sample : M1787-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-126D2-20210319

Quant Time: Mar 25 06:58:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 24 12:25:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	66944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	119254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	124951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	54068	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	56473	50.70	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.40%		
35) Dibromofluoromethane	5.464	113	42268	49.90	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.80%		
50) Toluene-d8	8.694	98	152920	50.66	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.32%		
62) 4-Bromofluorobenzene	11.118	95	61963	52.80	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.60%		
Target Compounds						
						Qvalue
16) Acetone	2.422	43	1876	4.45	ug/l #	73
24) 1,1-Dichloroethane	3.664	63	1759	1.23	ug/l #	81
27) cis-1,2-Dichloroethene	4.570	96	1246	1.45	ug/l	95
44) Trichloroethene	7.188	130	426496	487.08	ug/l	91
64) Tetrachloroethene	9.312	164	849	1.01	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032421\
Data File : VX021471.D
Acq On : 25 Mar 2021 06:06
Operator : JC/MD
Sample : M1787-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-126D2-20210319

Quant Time: Mar 25 06:58:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032421W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 24 12:25:26 2021
Response via : Initial Calibration

