

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034716.D
 Acq On : 24 Mar 2023 13:18
 Operator : JC/MD
 Sample : PB151645TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB151645TB

Quant Time: Mar 27 01:29:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	327067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	554558	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	501459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	229764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	244254	51.334	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.660%
35) Dibromofluoromethane	5.385	113	177627	48.513	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	8.647	98	672664	49.213	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.420%
62) 4-Bromofluorobenzene	11.079	95	257915	46.715	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.440%
Target Compounds						
16) Acetone	2.386	43	27049	12.068	ug/l	96
20) Methylene Chloride	2.788	84	14270	3.280	ug/l	96
37) Ethyl Acetate	4.709	43	64611	10.516	ug/l	99
43) Isopropyl Acetate	6.336	43	270057	25.947	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034716.D
Acq On : 24 Mar 2023 13:18
Operator : JC/MD
Sample : PB151645TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB151645TB

Quant Time: Mar 27 01:29:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

