

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034725.D
 Acq On : 24 Mar 2023 16:52
 Operator : JC/MD
 Sample : 02056-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 025-OUTFALL

Quant Time: Mar 27 01:33:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	300548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	509350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	453705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	222763	50.948	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.900%			
35) Dibromofluoromethane	5.391	113	161655	48.069	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.140%			
50) Toluene-d8	8.647	98	609298	48.533	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.060%			
62) 4-Bromofluorobenzene	11.079	95	235450	46.431	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.860%			
Target Compounds						
16) Acetone	2.386	43	20125	9.771	ug/l	94
52) Toluene	8.720	92	2334	0.249	ug/l	99
70) Styrene	10.659	104	117475	10.598	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034725.D
Acq On : 24 Mar 2023 16:52
Operator : JC/MD
Sample : 02056-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
025-OUTFALL

Quant Time: Mar 27 01:33:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

