

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034728.D
 Acq On : 24 Mar 2023 18:01
 Operator : JC/MD
 Sample : 02048-05DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4DL

Quant Time: Mar 27 01:34:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	310453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	523617	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	470010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	208033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	225059	49.831	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.660%
35) Dibromofluoromethane	5.385	113	163949	47.423	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.840%
50) Toluene-d8	8.647	98	621686	48.171	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.340%
62) 4-Bromofluorobenzene	11.079	95	240125	46.063	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.120%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	13408	6.302	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	500161	41.102	ug/l	96
20) Methylene Chloride	2.788	84	2327	0.564	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034728.D
Acq On : 24 Mar 2023 18:01
Operator : JC/MD
Sample : 02048-05DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4DL

Quant Time: Mar 27 01:34:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

