

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034729.D
 Acq On : 24 Mar 2023 18:25
 Operator : JC/MD
 Sample : 02048-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7DL

Quant Time: Mar 27 01:34:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	311784	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	531887	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	478264	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	230483	50.814	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.620%
35) Dibromofluoromethane	5.385	113	168951	48.110	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	8.647	98	635074	48.443	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.880%
62) 4-Bromofluorobenzene	11.079	95	245106	46.287	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.580%
Target Compounds						Qvalue
11) Tert butyl alcohol	2.977	59	167377	177.202	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	287380	23.515	ug/l	98
22) Diisopropyl ether	3.757	45	7013	0.532	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034729.D
Acq On : 24 Mar 2023 18:25
Operator : JC/MD
Sample : 02048-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7DL

Quant Time: Mar 27 01:34:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

