

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034732.D
 Acq On : 24 Mar 2023 19:34
 Operator : JC/MD
 Sample : 02080-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-4

Quant Time: Mar 27 01:35:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	312124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	527667	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	473232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	214199	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	221639	48.811	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.620%
35) Dibromofluoromethane	5.385	113	166632	47.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.660%
50) Toluene-d8	8.646	98	631532	48.558	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.079	95	241031	45.882	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034732.D
Acq On : 24 Mar 2023 19:34
Operator : JC/MD
Sample : 02080-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-4

Quant Time: Mar 27 01:35:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

