

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034733.D
 Acq On : 24 Mar 2023 19:57
 Operator : JC/MD
 Sample : 02080-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-5

Quant Time: Mar 27 01:35:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	321286	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	6.763	114	545628	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	485876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	221777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	228661	48.921	ug/l	0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.840%
35) Dibromofluoromethane	5.385	113	173522	48.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.340%
50) Toluene-d8	8.647	98	650518	48.372	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.079	95	251706	46.337	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.680%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	9111	4.138	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	1079531	85.722	ug/l	98
22) Diisopropyl ether	3.751	45	3460	0.254	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034733.D
Acq On : 24 Mar 2023 19:57
Operator : JC/MD
Sample : 02080-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-5

Quant Time: Mar 27 01:35:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

