

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
 Data File : VX034734.D
 Acq On : 24 Mar 2023 20:21
 Operator : JC/MD
 Sample : 02080-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-6

Quant Time: Mar 27 01:36:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	308179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	524229	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	469232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	210898	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	219612	48.984	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.385	113	165838	47.913	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	8.647	98	620033	47.987	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.980%
62) 4-Bromofluorobenzene	11.079	95	237298	45.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8734	4.136	ug/l	91
19) Methyl tert-butyl Ether	3.111	73	90027	7.453	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032423\
Data File : VX034734.D
Acq On : 24 Mar 2023 20:21
Operator : JC/MD
Sample : 02080-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-6

Quant Time: Mar 27 01:36:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

