

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045405.D
 Acq On : 24 Mar 2025 11:19
 Operator : JC/MD
 Sample : VX0324WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0324WBL01

Quant Time: Mar 25 01:29:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137486	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49419	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64463	57.958	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.920%
35) Dibromofluoromethane	5.379	113	49294	53.619	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.240%
50) Toluene-d8	8.647	98	168565	50.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	11.079	95	58097	52.604	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045405.D
Acq On : 24 Mar 2025 11:19
Operator : JC/MD
Sample : VX0324WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0324WBL01

Quant Time: Mar 25 01:29:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

