

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045408.D
 Acq On : 24 Mar 2025 12:34
 Operator : JC/MD
 Sample : Q1608-02DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE139D1-20250317DL

Quant Time: Mar 25 01:32:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130703	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	60881	58.394	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.780%			
35) Dibromofluoromethane	5.385	113	46554	53.267	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.540%			
50) Toluene-d8	8.647	98	162185	51.187	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.380%			
62) 4-Bromofluorobenzene	11.079	95	57281	54.557	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 109.120%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	1417	1.860	ug/l	98
12) 1,1-Dichloroethene	2.306	96	875	1.086	ug/l	95
44) Trichloroethene	7.123	130	46624	52.463	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045408.D
Acq On : 24 Mar 2025 12:34
Operator : JC/MD
Sample : Q1608-02DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE139D1-20250317DL

Quant Time: Mar 25 01:32:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

