

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045411.D
 Acq On : 24 Mar 2025 13:44
 Operator : JC/MD
 Sample : Q1608-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW11-MW01I-20250317

Quant Time: Mar 25 01:33:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135280	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61250	55.603	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.200%
35) Dibromofluoromethane	5.379	113	47388	52.387	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.780%
50) Toluene-d8	8.647	98	163534	49.867	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.079	95	55186	50.783	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.560%
Target Compounds						Qvalue
16) Acetone	2.386	43	969	1.896	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045411.D
Acq On : 24 Mar 2025 13:44
Operator : JC/MD
Sample : Q1608-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW11-MW01I-20250317

Quant Time: Mar 25 01:33:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

