

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045412.D
 Acq On : 24 Mar 2025 14:07
 Operator : JC/MD
 Sample : Q1608-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GM38-RW3-MW1-20250318

Quant Time: Mar 25 01:33:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	72484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64677	56.096	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.200%
35) Dibromofluoromethane	5.385	113	49839	52.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.220%
50) Toluene-d8	8.647	98	174748	50.403	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.800%
62) 4-Bromofluorobenzene	11.079	95	59982	52.210	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.420%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	905	1.692	ug/l	96
44) Trichloroethene	7.122	130	18390	18.911	ug/l	98
64) Tetrachloroethene	9.274	164	1290	1.564	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045412.D
Acq On : 24 Mar 2025 14:07
Operator : JC/MD
Sample : Q1608-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GM38-RW3-MW1-20250318

Quant Time: Mar 25 01:33:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

