

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045413.D
 Acq On : 24 Mar 2025 14:30
 Operator : JC/MD
 Sample : Q1608-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20250318

Quant Time: Mar 25 01:34:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	131180	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119040	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60693	57.436	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.880%	
35) Dibromofluoromethane	5.385	113	45614	52.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.000%	
50) Toluene-d8	8.647	98	160772	50.557	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.120%	
62) 4-Bromofluorobenzene	11.079	95	54900	52.099	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.200%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1312	1.569	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.319	101	9946	12.882	ug/l	96
12) 1,1-Dichloroethene	2.307	96	2198	2.693	ug/l	93
16) Acetone	2.392	43	779	1.589	ug/l #	70
27) cis-1,2-Dichloroethene	4.489	96	1634	1.641	ug/l	92
30) Chloroform	5.093	83	2448	1.488	ug/l	89
44) Trichloroethene	7.123	130	146962	164.765	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045413.D
Acq On : 24 Mar 2025 14:30
Operator : JC/MD
Sample : Q1608-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20250318

Quant Time: Mar 25 01:34:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

