

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045414.D
 Acq On : 24 Mar 2025 14:54
 Operator : JC/MD
 Sample : Q1608-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D2-20250318

Quant Time: Mar 25 01:34:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130401	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119535	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61019	57.340	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.680%			
35) Dibromofluoromethane	5.379	113	46462	53.285	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.560%			
50) Toluene-d8	8.647	98	159947	50.598	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.200%			
62) 4-Bromofluorobenzene	11.079	95	56165	53.618	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 107.240%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	1624	1.928	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.319	101	3818	4.910	ug/l	96
16) Acetone	2.380	43	879	1.780	ug/l	92
44) Trichloroethene	7.129	130	1028	1.159	ug/l	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045414.D
Acq On : 24 Mar 2025 14:54
Operator : JC/MD
Sample : Q1608-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D2-20250318

Quant Time: Mar 25 01:34:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

