

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045420.D
 Acq On : 24 Mar 2025 17:14
 Operator : JC/MD
 Sample : Q1608-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20250319

Quant Time: Mar 25 01:37:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.550	168	69702	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.757	114	134926	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.049	117	126257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.018	152	50122	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.958	65	62103	56.014	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.020%	
35) Dibromofluoromethane		5.379	113	47129	52.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.480%	
50) Toluene-d8		8.647	98	164753	50.370	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.740%	
62) 4-Bromofluorobenzene		11.079	95	57684	53.221	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.440%	
Target Compounds							
16) Acetone		2.380	43	1210	2.352	ug/l	99
44) Trichloroethene		7.123	130	5669	6.179	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045420.D
Acq On : 24 Mar 2025 17:14
Operator : JC/MD
Sample : Q1608-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20250319

Quant Time: Mar 25 01:37:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

