

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045422.D
 Acq On : 24 Mar 2025 18:00
 Operator : JC/MD
 Sample : Q1608-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT161S1-20250319

Quant Time: Mar 25 01:38:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59840	55.429	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.860%			
35) Dibromofluoromethane	5.379	113	45405	51.004	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.000%			
50) Toluene-d8	8.647	98	155997	48.335	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.680%			
62) 4-Bromofluorobenzene	11.079	95	54765	51.208	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.420%			
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.319	96	568	0.681	ug/l	98
16) Acetone	2.386	43	768	1.533	ug/l #	87
19) Methyl tert-butyl Ether	3.111	73	2141	0.765	ug/l	93
24) 1,1-Dichloroethane	3.611	63	4647	2.746	ug/l	98
44) Trichloroethene	7.129	130	5695	6.291	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045422.D
Acq On : 24 Mar 2025 18:00
Operator : JC/MD
Sample : Q1608-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT161S1-20250319

Quant Time: Mar 25 01:38:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

