

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
 Data File : VX045427.D
 Acq On : 24 Mar 2025 19:56
 Operator : JC/MD
 Sample : Q1608-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D3-20250319

Quant Time: Mar 25 01:40:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	66010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128913	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58266	55.492	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.980%		
35) Dibromofluoromethane	5.379	113	44214	51.292	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.580%		
50) Toluene-d8	8.647	98	153598	49.150	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 98.300%		
62) 4-Bromofluorobenzene	11.079	95	52579	50.774	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 101.540%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	843	1.730	ug/l #	22
29) Tetrahydrofuran	5.026	42	2619	5.318	ug/l #	72
44) Trichloroethene	7.141	130	580	0.662	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX032425\
Data File : VX045427.D
Acq On : 24 Mar 2025 19:56
Operator : JC/MD
Sample : Q1608-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D3-20250319

Quant Time: Mar 25 01:40:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

