

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
 Data File : VX008429.D
 Acq On : 25 Mar 2019 22:56
 Operator : JC/SP
 Sample : K1969-09RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-DUP06-20190316RE

Quant Time: Mar 26 07:00:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 04:58:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	236115	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	400669	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	346059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144281	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	168893	47.18	ug/l	0.00
Spiked Amount			Recovery	=	94.36%	
35) Dibromofluoromethane	5.50	113	125269	47.95	ug/l	0.00
Spiked Amount			Recovery	=	95.90%	
50) Toluene-d8	8.71	98	503404	48.19	ug/l	0.00
Spiked Amount			Recovery	=	96.38%	
62) 4-Bromofluorobenzene	11.13	95	172375	44.53	ug/l	0.00
Spiked Amount			Recovery	=	89.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	5428	2.90	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032519\
Data File : VX008429.D
Acq On : 25 Mar 2019 22:56
Operator : JC/SP
Sample : K1969-09RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-DUP06-20190316RE

Quant Time: Mar 26 07:00:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 04:58:03 2019
Response via : Initial Calibration

