

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
 Data File : VX008432.D
 Acq On : 26 Mar 2019 00:05
 Operator : JC/SP
 Sample : K1969-05RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW301I-20190316RE

Quant Time: Mar 26 07:05:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 04:58:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	227788	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	389973	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	335893	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137708	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	163691	47.40	ug/l	0.00
Spiked Amount			Recovery	=	94.80%	
35) Dibromofluoromethane	5.50	113	120363	47.33	ug/l	0.00
Spiked Amount			Recovery	=	94.66%	
50) Toluene-d8	8.71	98	486741	47.87	ug/l	0.00
Spiked Amount			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	11.13	95	163331	43.35	ug/l	0.00
Spiked Amount			Recovery	=	86.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	4843	2.68	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032519\
Data File : VX008432.D
Acq On : 26 Mar 2019 00:05
Operator : JC/SP
Sample : K1969-05RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW301I-20190316RE

Quant Time: Mar 26 07:05:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 04:58:03 2019
Response via : Initial Calibration

