

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX032519\
 Data File : VX008437.D
 Acq On : 26 Mar 2019 02:02
 Operator : JC/SP
 Sample : K2059-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Mar 26 07:16:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 26 04:58:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	405081	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	349467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145681	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	171772	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
35) Dibromofluoromethane	5.50	113	125025	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
50) Toluene-d8	8.71	98	508615	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
62) 4-Bromofluorobenzene	11.13	95	172545	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
Target Compounds						
16) Acetone	2.44	43	5437	2.891	ug/l	96
25) 2-Butanone	4.68	43	9311	3.088	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX032519\
Data File : VX008437.D
Acq On : 26 Mar 2019 02:02
Operator : JC/SP
Sample : K2059-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Mar 26 07:16:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032519W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 26 04:58:03 2019
Response via : Initial Calibration

